

नाशिक महानगरपालिका, नाशिक

सार्वजनिक आरोग्य विभाग

(सन २०२४-२५)

ई - कोटेशन नोटीस क्र.०१

दि.२६/०७/२०२४

नाशिक महानगरपालिका सार्वजनिक विभागांतर्गत रुग्णालयातील डेग्यू संशयीत रुग्णांच्या तपासणी करिता आवश्यक किट व अनुषंगिक साहित्य इ.सन २०२४-२५ वर्षासाठी आवश्यकतेनुसार खरेदी कामी जाहीर ई-कोटेशन द्वारे दर मागविणेत येत आहे.

१. कोटेशन फॉर्म मिळण्याचा कालावधी : दि.२६/०७/२०२४ ते.
दि.०२/०८/२०२४
२. कोटेशन फॉर्म भरून जमा करणेची अंतिम तारीख : दि.०२/०८/२०२४
३. कोटेशन फॉर्म मिळण्याचे ठिकाण व वेळ : वैद्यकिय (आरोग्य) अधिकारी कार्यालय,
राजीव गांधी भवन, ३रा मजला, शरणपुर
रोड, नाशिक.
वेळ : सकाळी १०:०० ते ४:०० पर्यंत.
संकेतस्थळ - www.nmc.gov.in

कोटेशन फॉर्म पुर्ण भरून बंद लिफाफ्यात मा.वैद्यकिय (आरोग्य) अधिकारी यांचे नावे मनपा राजीव गांधी भवन, सार्वजनिक आरोग्य विभाग येथे कार्यालयीन वेळेत देण्यात यावे. अपुर्ण कोटेशन स्विकारले जाणार नाही. कोटेशन फॉर्म सोबत पॅन कार्ड व जी.एस.टी.रजि.प्रमाणपत्र जोडणे अनिवार्य आहे. कोटेशन फॉर्म स्विकारण्याचा अथवा नाकारण्याचा हक्क राखून ठेवण्यात आला आहे.

सादर करावयाची कागदपत्रे :

- १) दरपत्रक (कोटेशन फॉर्म)
- २) GST प्रमाणपत्र सत्यप्रत
- ३) PAN कार्ड सत्यप्रत

Sd/-

वैद्यकिय (आरोग्य) अधिकारी
नाशिक महानगरपालिका, नाशिक.



नाशिक महानगरपालिका, नाशिक
कोटेशन फॉर्म

दि.२६/०७/२०२४

निम्नलिखित तपशीलाप्रमाणे नाशिक महानगरपालिका सार्वजनिक विभागांतर्गत रुग्णालयातील डेंग्यू संशयीत रुग्णांच्या तपासणी करिता आवश्यक किट व अनुषंगिक साहित्य इ. इ.सन २०२४-२५ वर्षासाठी आवश्यकतेनुसार खरेदी करणेकामी बाजारभाव (Quotation) बोलविणेत येत आहेत.तरी त्याबाबतचे दर बंद पाकिटातून दिनांक ०२/०८/२०२४ पर्यंत इकडील कार्यालयास पाठविणेचे करावे.

सदरचे दर पुरवठादारास सन २०२४-२५ या वर्षासाठी बंधनकारक राहतील. दर हे सर्व करांसहीत व सोबत नमूद तांत्रिक विनिर्देशानुसार (specification) सादर करावेत.

अ.क्र.	किटचे नाव	पॅकिंग	दर (सर्व करांसह)
1	Dengue NS-1 Antigen Rapid Test	1 Test	
2	Rapid Malaria card kit (PF/PV)	1 Test	
3	Glass slide (microscopic slide)	1 x 50 nos	
4	Plastic Slide Box (for 50 slides)	1 No	
5	Pricker / Lanset	1 x 100 nos	
6	Rapid malaria stain kit (Filled A&B)	2 x 100 ml	

टिप- वरील सर्व बाबी करीता प्रती टेस्ट, प्रती पॅक, प्रती नग, प्रती 100 .लि.
नुसार सर्व करांसह दर नोंदवावे.

पुरवठादाराची सही व शिक्का

Technical Specifications

<p>1) <u>Dengue NS-1 Antigen Rapid Test</u> :</p> <ol style="list-style-type: none">1.Rapid visual test for Dengue NS 1 Antigen and IgM & IgG Antibodies detection2. Positive and negative controls provided with each pack of kit3. Diagnosis of both Primary & Secondary Infection4. Detection of all 4 serotypes of Dengue Virus5. Highly Sensitive & Highly Specific6. Long Shelf Life 18 months at 2-30°C
<p>2) <u>Rapid Malaria card kit (PF/PV)</u> :</p> <ol style="list-style-type: none">1. Antigen of P.Falciparum (Pf) and P.vivax (Pv) from human serum or plasma or whole blood2. Positive and negative controls provided with each pack of kit3. Each test card/cassette with desiccant individually packed in a hermetically sealed and non-permeable pouch4. The test kit packed in such a way that there is provision to conduct single test at a time
<p>3) <u>Glass-slide Box (microscopic)</u> :</p> <ol style="list-style-type: none">1.Thickness of slide should be 0.95mm – 1.15mm.2.Length and width of the glass slide should be 75mm x 25mm.3.Edge of the slide shall be ground to become free from sharp edges and chips.4.material of this glass should be transparent and colourless.
<p>4) <u>Plastic Slide Box (for 50 slides)</u> :</p> <ol style="list-style-type: none">1. The slide box capacity should be 50 atleast and the material should be polystyrene.2. Working Temperature should be -80Deg C - 121Deg C
<p>5) <u>Pricker/ Lanset</u> :</p> <ol style="list-style-type: none">1. Material should be Medical grade Stainless Steel and Polymer.2. Shelf life should be 36 months from the date of manufacture.3. Needle size should be 28 gauge
<p>6) <u>Rapid malaria stain kit (Filled A&B)</u> :</p> <ol style="list-style-type: none">1. All reagents should be ready to use2. All reagents should be stable at RT for 3 years.3. It should be Polychromated methylene blue and Eosin stains specifically to basophilic and acidophilic cellular elements to demonstrate blood cells and hemoparasites4. It should be able to discovering malarial parasites

पुरवठादाराची सही व शिक्का