



नाशिक महानगरपालिका, नाशिक

कोटेशन नोटीस क्र. ०९

नाशिक महानगरपालिका सार्वजनिक आरोग्य विभागांतर्गत पाथर्डी गाव UPHC करीता आवश्यक साहित्य व फर्निचर खरेदीकामी कोटेशन मागविणेत येत आहे.

१.कोटेशन फॉर्म मिळण्याचा कालावधी : दि.०५/१२/२०२३ ते दि.१२/१२/२०२३

२.कोटेशन फॉर्म भरुन जमा करणेची अंतिम तारीख : दि. १२/१२/२०२३

३.कोटेशन फॉर्म मिळण्याचे ठिकाण व वेळ : वैद्यकीय(आरोग्य)अधिकारी
कार्यालय,राजीव गांधी भवन,
३ रा मजला,शरणपुर रोड.नाशिक
वेळ सकाळी १०:०० ते ४:०० व
संकेतस्थळ www.nmc.gov.in

कोटेशन फॉर्म पुर्ण भरुन बंद लिफाफ्यात मा.वैद्यकीय(आरोग्य)अधिकारी यांचे नावे मनपा राजीव गांधी भवन, ३ रा मजला, सार्वजनिक आरोग्य विभाग येथे कार्यालयीन वेळेत देण्यात यावे, अपुर्ण कोटेशन स्विकारले जाणार नाही.कोटेशन फॉर्म स्विकारण्याचा अथवा नाकारण्याचा हक्क राखुन ठेवण्यात आला आहे.

सादर करावयची कागदपत्रे

१) दरपत्रक(कोटेशन फॉर्म)

२) GST प्रमाणपत्र सत्यप्रत

३) PAN कार्ड सत्यप्रत

Sd/-

वैद्यकीय(आरोग्य)अधिकारी
नाशिक महानगरपालिका, नाशिक



नाशिक महानगरपालिका नाशिक
कोटेशन फॉर्म

दि. / /२०२३

निम्नलिखित तपशिला प्रमाणे खालील तक्त्यात दर्शविले नुसार नाशिक मनपा सार्वजनिक आरोग्य विभागांतर्गत पाथर्डी गाव UPHC करीता आवश्यक साहित्य व फर्निचर खरेदी करावयाचे असलेने ते पुरविणेबाबत बाजारभाव (Quotation) मागविणेत येत आहेत. तरी त्याबाबतचे दर (सर्व करांसहित) बंद पाकिटातुन दि. १२/१२/२०२३ पर्यंत इकडील कार्यालयात पाठविण्याचे करावे.

सदचे दर पुरवठादार यांस पुरवठा आदेशाच्या दिनांका पासुन पुढील एक वर्ष कालावधी करीता बंधनकारक राहिल.सदर मालाचा पुरवठा, पुरवठा आदेश मिळाल्यापासुन त्वरीत करावा लागेल. मालाचे दर हे सर्व करांसहीत देण्यात यावेत.

Sr No	Name Of Item	Req.Qty	Rate/Nos (Inclusive All Taxes)	Total (Inclusive All Taxes)
1	IUCD insertion Kit	1		
2	Spong holding forceps 10"	2		
3	Vulsellum utrine forceps	2		
4	Kidney Tray	2		
5	Touch With 3 dry Cell	2		
6	Bowl for Antiseptic solution for soaking cotton swabs	2		
7	Tray Containing chlorine solution for keeping soiled Instrument	1		
8	Head light	1		
9	B.P.Appratus (Digital)	2		
10	Stethoscope Adult	2		
11	General Instrument Set	1		
12	Examination Table (with mattress)	2		
13	Writing tables	5		
14	Dr.Chair	3		
15	Plastic Chair	10		
16	Armless chairs	10		
17	Full size steel almirah	5		
18	Bench for waiting area	10		
19	Stretcher on trolley	2		
20	Foot step(Double step)	2		
21	Stool (S.S.Top)	5		
22	Plastic Bucket 20 ltr	4		
23	Plastic Mugs	4		
24	Bed sheets	12		
25	Pillows With Covers	12		
26	Blankets(Red)	6		
27	Towels	4		
28	Curtain With Rod with fitting for 26 Window (1Set)	1		
29	Dustbin	5		
30	Tongue Depressor	2		
31	Thermometer Clinical	2		
32	Needle Destroyer (Manual)	2		
33	Instrument Trolley	1		
34	IV Stand	5		
Total				

Sr. No	Name of Items & Furniture	Specification
1	IUCD insertion Kit	Contain 1) Sponge holding forceps 10" -1Nos 2) Cheatle's forceps.8" -1Nos 3) Vulsellum/Allis forceps.10" -1Nos 4)Anterior Vaginal Wall Retractor 6" - 1Nos 5) Sims Bivalve Speculum, -1 Nos 6) Uterine Sound length 10" -1Nos 7) Bowl SS. – 2 Nos 8) Curved Scissors (Blade Size 765 mm) -1Nos 9)Long artery straight forceps (for IUCD removal). 10" -1Nos 10) Kidney tray 10" -1Nos 11) ISO certified.
2	Sponge holding forceps	1) Sponge Holding Forceps 10" 2) Stainless Steel Quality. 3) It should be CE and ISO certified. 4) Easy to use.
3	Vulsellum utrine forceps	1)Vulsellum uterine Forcep 10" 2) Stainless Steel Quality. 3)It should be CE and ISO certified. 4)Easy to use.
4	Kidney Tray	1)It should be made up of Stainless steel. 2) It should not have sharp edges at curve. 3) It should be CE and ISO certified. 4) Dimensions: 10"
5	Tourch With 3 dry Cell	Light weighted, sturdy, similar to pen torch, weather resistant, durable, power consumption 2W - 5 W with appropriate lumens.
6	Bowl for Antiseptic solution for soaking cotton swabs	1) Dimensions of Bowl 12 ×12× 3.5 cm (L×W×H) 2) Material: Stainless steel 3) 1 Year Warranty
7	Tray Containing chlorine solution for keeping soiled Instrument	1) It should be made up of Stainless steel. 2) It should not have sharp edges at curve. 3) Should have led to cover the tray. 4) Easy to clean. 5) It should be CE and ISO certified. 6) Dimensions: 12"x10"
8	Head light	1) Color Temperature (Degree Kelvin) 2700-3000 degree K 2) Minimum LED Life (Hours) 10000 3) Charging time 2 Hours 4) Color rendering index (CRI) - 85 5) Run time (Minures) 300 Product complies with IEC 60601-1 and IEC60601-2-41 Yes
9	B.P.Appratus (Digital)	1) Fully Automatic inflation and deflation 2) Hypertension indicator (Above 140 Systolic/ 90 diastolic) 3) Memory for at least 5 sets of last readings 4) Display Digital. Machine should be portable 5)BP Measurement Range - (0- 299) mm of Hg 6) Pulse recorder - 40 - 180 beats/ min 7)Accuracy: Pressure + 4 mm of Hg Pulse reading + 5% of reading 8) Inflation : Electric Pump 9)Deflation : Automatic pressure relax value 10) Power Supply : 4 AA Battery 1.5 V (Rechargeable batteries +charger) 11) Battery life : Approximately 300 measurements 12) Operating Temperature - 10 to 40 degree Celsius 13) Cuff Material : (Double stitched) Nylon or Polyester 14) Medium Size cuff Qty- 02Nos:, Large Size cuff Qty- 01No.with arterial indicator 15) 2 years Manufacturer warranty 16) Laminated pictorial user Manual 17) ISO certified
10	Stethoscope Adult	1) Features:-Tunable

Sr. No	Name of Items & Furniture	Specification
		2) Chest pieces material :-Stainless steel 3) Types :-Acoustic 4)Types by practice :-physician, pediatrician 5) It should be CE Certified. 6) 1 Year Warranty
11	General Instrument Set	Contain 1) Plain forceps -1Nos 2) Tooth forceps-1Nos 3) Allies forceps size 6" -1Nos 4) Allies forceps size 8" -1Nos 5) Artery forceps size 6" -1Nos 6) Artery forceps size 8" -1Nos 7) Sponge holding forceps 6"-1Nos 8) Sponge holding forceps 8"-1Nos 9) Stitch removal scissor-1Nos 10) Blunt scissor-1Nos 11) Knee Hammer- 1Nos-1Nos 12) Stainless Steel Quality. 13) It should be CE and ISO certified. 14) Easy to use.
12	Examination Table (with mattress)	1) 3 drawer, 3 cupboard, BP apparatus stand & foot step stool 2) Three Cabinet, two Drawers with Lock, writing pad, Attached foot step, fix ply mattress with Raxine covered. 3) Height should be 23"X72"X33"H 4) 1 Year warranty 5) It should be ISO-9001 certified.
13	Writing tables	1)Dimensions: 48"X24"X30" H. 2)Powder coated table one side 3 drover with lock, One side Cabinet. 3) 18 mm wooden top with sunmica. 4) It should be ISO certified. 5) 1 year warranty.
14	Doctor Chair	1) Superior Lumbar Support 2) Quality Leatherette and Gas-Lift Mechanism. 3) Soft Cushioned seat & padded Arms Ensures full comfort. 4) Dimensions 95 X 45 X 60 cm. 5) Pneumatic Seat-Height Adjustment, 5) 360 degree Swivel & Smooth Rolling Casters. 6) Warranty 1 year. 7) ISO certified.
15	Plastic Chair	1) Single piece Plastic molded with arm, without cushion. 2) Plastic material- Polypropylene 3) Seat width 365 mm (+/- 5mm) 4) Seat depth 360 mm (+/- 5mm) 5) Chair height 500mm (+/-10 mm) 6) Color Off White 5)Warranty 1 Year
16	Armless Chairs	1)Non revolving And non tilting chair Without armrest 2)Seat – Padded 3)Backrest- Padded 4)Frame Material configuration one piece 5)Frame material-Steel tube(ERW quality) Conforming to IS1161 (with latest Amendment) 6)Seat material-Molded Plywood 7) Seat material Thickness 12 mm 8) Seat upholstered with Fabric 9) GSM of seat cover 220 GSM (+/- 3 GSM) 10) Backrest material-Molded Plywood 11) Density of Polyurethane foam used in seat& Backrest 35 (+2 Kg/Cub Meter) 12)Seat width 540 mm (+/- 5mm) 13)Seat depth 450 mm (+/- 5mm) 15)Seat height from floor Surface 475mm (+/- 5mm)

Sr. No	Name of Items & Furniture	Specification
		16) Backrest width 410 mm (+/- 5mm) 17) Backrest height from Seat 360 mm (+/- 5mm) 18) Black color epoxy paint.
17	Full size steel almirah	1) It should be MS Powder coated with 5 selves 2) Dimensions: 36"X15"X78"H 3) 1 Year warranty 4) It should be ISO certified 9001.
18	Bench for waiting area	1) It should be made up of Wooden top & with powder coated MS metal Frame. 2) Size: 72 (L)X 24(W) X18(H) inches 3) Maximum Weight Recommendation 300Kg to 350 Kg 4) 1 Year warranty 5) ISO certified
19	Stretcher on trolley	1) Stretcher on trolley SS top 4" caster powder coated trolley, 2) Stretcher top SS, trolley move on 4" castor wheel. 3) Dimensions: 22"(W)X78"9(L)X32" (H) 4) It should be ISO certified
20	Foot step (Double step)	1) Foot step double 2) Powder coated 3) Dimensions: 15"X16"X15"H 4) It should be ISO certified.
21	Stool (S.S.Top)	1) Bed Side Stool SS Top: 10"X10"X18" 2) With powder coated legs 3) 1 Year warranty 4) It should be ISO certified 9001
22	Plastic Bucket 20 ltr	1) It should be made up of Virgin Plastic 2) It should be Strong enough to lift with handle 3) Handle should be strong and durable 4) Warranty 1 years. 5) ISO certified
23	Plastic Mugs	1) It should be made up of Virgin Plastic 2) It should be Strong enough to lift with handle 3) Handle should be strong and durable 4) Warranty 1 years. 5) ISO certified.
24	Bed sheets	1) It Should be made up of cotton. 2) It should be white in color. 3) Size 60" X 72" 4) Warranty 1 years. 5) ISO certified
25	Pillows With Covers	1) It Should be made up of cotton. 2) It should be white in color. 3) Size 18" X 24" 4) Warranty 1 years. 5) ISO certified.
26	Blankets(Red)	1) It Should be made up of wool/Polyamide/ Polyester. 2) It should be RED in color. 3) Size 60" X 90" 4) It should weight 1.5 Kg 4) Warranty 1 years. 5) ISO certified
27	Towels	1) It Should be made up of Cotton. 2) Size 18" X 24" 3) Warranty 1 years. 4) ISO certified
28	Curtain with Rods with Fitting.	1) 4.6 ft X 5 ft=15 Window 2) 4.6 ftX4ft= 4 Window 3) 4 ftX4ft= 4 Window 4) 4.5 ftX5 ft=3 Window
29	Dustbin	1) It should be made up of Virgin Plastic 2) It should be Strong enough to lift 3) It should have Lid to cover dust bin 4) It should be 20 lit in capacity.

Sr. No	Name of Items & Furniture	Specification
		4) Warranty 1 years. 5) ISO certified
30	Tongue Depressor	1) It should be made up of Stainless Steel 2) It should not have sharp edges 3) It should be autoclavable and reusable 4) Warranty 1 years. 5) ISO certified
31	Thermometer Clinical	1) Range of temperature measurement 32.0C- 42.0C (89.60F-109.40F) 2) Can be calibrated in both centigrade and Fahrenheit, but if only one option is available, then Fahrenheit is preferable. 3) Buzzer signal function. 4) Takes 60-90 seconds to measure temperature. 5) Can be used in the armpit/maxilla, orally and rectally. 6) Accuracy of temperature ± 0.1 deg C and ± 0.2 F. 7) User's interface LCD display. 8) Battery operated 9) It Should be CE/ ISO Certified
32	Needle Destroyer (Manual)	1) Dimensions (Minimum) 165mm x 160mm x 180mm(LxWxH) (Excluding Handle) 2) Expected Life 1,00,000 Cuts 3) Material Body Stainless Steel & Alloy Steel 4) Sharp Container capacity 1500 ml compatible to cutter 5) Mobility, Portability Portable 6) Power Requirements NIL 7) Accessories (Mandatory) Sharp Containers 8) Warranty 1 years. 9) ISO certified.
33	Instrument Trolley	1) Instrument trolley regfull SS, 2) SS Top, trolley move on dia 100 mm castor Wheels. 3) Dimension 18"x17"x32" 4) It should be ISO certified.
34	IV Stand	1) IV STAND POWDER COATED WITH SS ROD 2) Height should be 82" 3) 1 Year warranty. 4) It should be ISO certified 9001.

पुरवठादाराची सही व शिक्का