



नाशिक महानगरपालिका,नाशिक

कोटेशन नोटीस क्र. ०४

नाशिक महानगरपालिका सार्वजनिक आरोग्य विभागांतर्गत मोरवाडी प्राथमिक आरोग्य केंद्र,मनपा (स्वामी समर्थ रुग्णालय) येथे खासदार स्थानिक क्षेत्र विकास कार्यक्रम अंतर्गत लागणारे रुग्णालयीन साहित्य १) Incubator (Paediatric Radiant Warmer) २) Foetal Doppler ३) Pulse Oximeter ४) Multiparamonitor व ५) OT Table (Hydraulic Gynecological Table) खरेदी करणेकामी कोटेशन मागविणेत येत आहे.

१.कोटेशन फॉर्म मिळण्याचा कालावधी : दि.२३/०९/२०२२ ते दि.३०/०९/२०२२

२.कोटेशन फॉर्म भरून जमा करणेची अंतिम तारीख : दि.०३/१०/२०२२

३.कोटेशन फॉर्म मिळण्याचे ठिकाण व वेळ : वैद्यकिय(आरोग्य)अधिकारी कार्यालय,राजीव गांधी भवन, ३ रा मजला,शरणपुर रोड.नाशिक वेळ सकाळी १०:०० ते ४:००

कोटेशन फॉर्म पुर्ण भरून बंद लिफाफ्यात मा .वैद्यकिय(आरोग्य)अधिकारी यांचे नावे मनपा राजीव गांधी भवन, सार्वजनिक आरोग्य विभाग येथे कार्यालयीन वेळेत देण्यात यावे , सोबत PAN व GST चे छायांकित प्रमाणपत्र जोडावे. अपुर्ण कोटेशन स्विकारले जाणार नाही .कोटेशन फॉर्म स्विकारण्याचा अथवा नाकारण्याचा हक्क राखून ठेवण्यात आला आहे.

Sd/-

वैद्यकीय (आरोग्य)अधिकारी
नाशिक महानगरपालिका, नाशिक



नाशिक महानगरपालिका नाशिक कोटेशन फॉर्म

दि. / /२०२२

निम्नलिखित तपशिला प्रमाणे खालील तक्त्यात दर्शविलेले रुग्णालयीन साहित्य खरेदी करणे असलेने ते पुरविणेबाबत बाजारभाव (Quotation) मागविणेत येत आहेत. तरी त्याबाबतचे दर (सर्व करांसहित) बंद पाकिटातुन दि. ०३/१०/२०२२ पर्यंत इकडील कार्यालयात पाठविण्याचे करावे.

सदचे दर पुरवठादार यास पुरवठा आदेशाच्या दिनांका पासुन पुढील एक वर्ष कालावधी करीता बंधनकारक राहिल . सदर मालाचा पुरवठा आदेश मिळाल्यापासुन पंधरा दिवसात करावा लागेल.मालाचे दर हे सर्व करांसहीत देण्यात यावेत.

अ. क्र.	साहित्याचे नाव	संख्या	अंदाजीत दर/ प्रती नग (सर्व करांसहित)	एकुण र.रु. (सर्व करांसहित)
1.	Incubator (Paediatric Radiant Warmer)	3		
2.	Foetal Doppler	4		
3.	Pulse Oximeter	2		
4.	Multiparamonitor	1		
5.	OT Table (Hydraulic Gynecological Table)	1		

Sr. No.	Name of Equipment	Specification
1.	Incubator (Paediatric Radiant Warmer)	<p>Micro computer Temperature controller:</p> <ul style="list-style-type: none">• PID (proportionate integrated derivative) based micro processor technology assures accurate maintains of set temperature with internal accuracy of 0.01°C• User friendly on-board temperature calibration with feather touches keys.• Power Source: 230V AC, ±10%, 50Hz.• Light Source: LED light to examine babies.• Heater Capacity: 600 Watts.• Power Consumption: 800 Watts at 100% Heater Power.• Product should be CE Mark.• 1 YEAR WARRANTY.
2.	Fetal Doppler	<ul style="list-style-type: none">• Ultrasound Frequency: 2MHz• Ultrasound intensity <10mW/cm²• Power Supply DC Ni-Mh rechargeable battery AC 220/110V, 50/60Hz• Display 45mmx25mm LCD• FHR Measuring range: 50-240bpm• FHR resolution: 1 bpm• FHR accuracy: ±1 bpm• Power consumption <1 W• Dimension: 135mmx95mmx35mm• Weight 500 gm• Product should be CE Mark.• 1 YEAR WARRANTY
3.	Pulse Oximeter (Finger Tip Pulse Oxymeter)	<ul style="list-style-type: none">• SpO₂ measurement : Transducer: Dual-wavelength LED sensor with Wavelength: Red Light: 663nm, infrared light: 890nm. Maximal average optical output power: <52mW• SpO₂ display range: 35%-100%• SpO₂ measuring accuracy: ≤2% for SpO₂ from 70% to 100%• Pluse Rate Measurement : PR display range: 30bpm-240bpm

		<p>PR measuring accuracy: ± 2bpm or $\pm 2\%$ (whichever is greater)</p> <ul style="list-style-type: none"> • Perfusion Index (PI) Display range: 0%-20% • Preset over- limits SpO2 low limit 90 % Pluse Rate: high Limit: 120bpm Low limit 50bpm • Over-limit Indication settings for PC-60c1, N.N.W: SpO2: low limit setting range: 85%-95% Default setting :85% Pluse Rate: Low limit setting range 30-60bpm; High limit setting range: 100-240bpm Default setting High:120bpm, Low: 50bpm • Power Supply requirement : 2xLR03 (AAA) alkaline batteries Supply voltage 3.0VDC Operating current: ≤ 40Ma Product Should be CE Mark & FDA APPORVED AND WITH 1 YEAR WARRANTY
4.	Multipara-- -monitor	<ul style="list-style-type: none"> • 12.1 TFT color LCD, multi-language interface. Should measure adult & Pediatric SPOZ, NIBP, RR, ECG and Temp. Fan less design, quiet, energy-saving and clean, which reduces the possibility of cross-infection. • All-round monitor for adult, pediatric and neonate with standard interface, oxygen Graph, trend graph, big character interface and view bed convenient observe. • Finish all operations by keys and knobs • Maximum 8-channel waveform synchronous display. • Display 7-lead ECG waveform on one screen, cascade ECG waveform display. • Adopt digital Spo2 technology anti-motion and anti-ambient light interference, and Measurement can be performed under the circumstance of weak filling. • Heart rate variability (HRV) analysis function • NIBP measurement mode: Manual/AUTO/STAT, storage for 4800 group NIBP data. • Review for 71 alarm events of all parameters and 60 arrhythmia alarm events. Drug concentration calculation and titration table functions. • One-touch printing of trend graph • Connect to central Monitoring System by 3G, WI FI or wired mode. • Product should be CE mark. • 1 YEAR WARRANTY. ACCESSORIES Warranty -6 MONTH WARRANTY.
5.	OT Table (Hydraulic Gynecologi cal Table)	<ul style="list-style-type: none"> • Length of Table : 1600 mm • Width of Table Top : 550 mm • Minimum Height : 750 mm • Maximum Height : 1000 mm • Trendeleburg : 25° • Lateral Tilt : $\pm 20^\circ$ • Back Section : $\pm 60^\circ$ • Head Section : $\pm 60^\circ$ • Leg section : $0^\circ - 90^\circ$ • Standard Accessories:- Shoulder support Pad, Knee Crutches, Lateral Support pad, Arm Boards Pad, Water proof mattress, Anesthetic Screen. Suitable for Obstetrics Operating room table perfect for all kinds of Gynecological operations. Good performance and easy to clean. Product should be CE Mark. 2 YEAR WARRANTY. Major Part of the Table is made of form SS 304 steel Grade to avoid Rust.